

Guest Editorial

Body Sensor Networks: From Theory to Emerging Applications

Abstract—The use of sensor networks for healthcare, well-being, and working in extreme environments has long roots in the engineering sector in medicine and biology community. With the maturity of wireless sensor networks, body area networks (BANs), and wireless BANs (WBANs), recent efforts in promoting the concept of body sensor networks (BSNs) aim to move beyond sensor connectivity to adopt a system-level approach to address issues related to biosensor design, interfacing, and embodiment, as well as ultralow-power processing/communication, power scavenging, autonomic sensing, data mining, inferencing, and integrated wireless sensor microsystems. As a result, the system architecture based on WBAN and BSN is becoming a widely accepted method of organization for ambulatory and ubiquitous monitoring systems. This editorial paper presents a snapshot of the current research and emerging applications and addresses some of the challenges and implementation issues.

Index Terms—Body sensor networks (BSNs), sensors, ubiquitous systems, wireless body area networks (WBANs).

I. INTRODUCTION

THIS SPECIAL issue marks almost ten years from the first implementation of wireless body area networks (WBANs) [1], though at that time, the prototype was named under personal area network (PAN), a term that was originated from Zimmerman [2] and further developed by IEEE P802.15 Working Group [3]. Soon after, the notation of BAN emerged. A group from Philips was among the first to use BAN instead of PAN and has listed distinct features that should be incorporated into the two types of networks [4]. As the applications of PAN or BAN extended from connecting personal electronic consumer goods for the sake of convenience to the user to medical and healthcare applications, BAN has become increasingly popular and is found to be a key element in the infrastructure for patient-centered medical applications [5].

The launch of projects, such as MobiHealth [6], further accelerates the development of BAN and m-Health service platform. IEEE T-ITB published a Special Section on m-Health five years ago [7] and the section turns out to be well recognized, as reflected in the number of citations.

From a system perspective, the concept of body sensor networks (BSNs) [8] moves beyond sensor connectivity, with specific focuses on ultralow-power processing/communication, power scavenging, autonomic sensing, data mining, distributed inferencing, intelligent on-node processing, and integrated wireless sensor microsystems. The primary motivation of BSN re-

search is to provide long-term continuous sensing without activity restriction and behavior modification.

In practice, a collection of sensors on a user can be integrated using wired, wireless, and biochannels or a combination of these approaches. Wired sensor networks are typical for smart clothes that integrate both sensors and sensor interconnections, and achieve ultimate power efficiency. Wireless sensor networks (WSNs) can cover sensors spanning the whole body. They are particularly important for linking with ambient sensors and implantable sensors, such as pacemakers and deep brain stimulators [9]. Biochannels serve as a unique secured means of communication [10], where the human body is used to transmit either exogenous or endogenous information. Since each approach has its pros and cons, a combination of these approaches should be used according to specific applications, i.e., formation of a hybrid network [11].

Recent technological developments have enabled sensor miniaturization, power-efficient design and improved biocompatibility. Issues related to system integration, low-power sensor interface, and optimization of wireless communication channels are active research fields, as presented in this special issue. Other topics related to quality of service, security, multi-sensor data fusion, decision support, and technological scaling are also important for practical applications of BSNs. The purpose of this special issue is to present recent trends and transition of BSN technology from theoretical concepts to emerging applications.

II. EMERGING TECHNOLOGIES AND APPLICATIONS

Earlier prototypes of BSNs tend to use multichip solutions manufactured from off-the-shelf components. Recently, the first system-on-chip (SoC) has been developed specifically for wireless BSNs to monitor vital signs. The SoC integrates a transceiver, hardware media access control (MAC) protocol, microprocessor, IO peripherals, memories, A-D converter, and custom sensor interfaces. The chip can be interfaced with antenna and battery as a single-chip wearable patch [12].

Thus far, long-term and sustainable power supply remains a great challenge to BSNs. New trends in very large-scale integration (VLSI) technology and laboratory prototypes developed by MIT and Texas Instruments promise to decrease processor power consumption 10–20 times and power supply voltage to less than 500 mV in the next five years [13]. While significant effort has been directed to power-efficient processor design and communication protocols, extensive studies have also been carried out in establishing practical techniques for power

scavenging from the external environment or the human body [14]. One potential direction is the development of biofuel cells, where biocatalysts are used for converting chemical energy of a fuel such as glucose into electrical energy [15].

Parallel to these developments, new materials and technologies have also been developed for innovative sensor embodiment to achieve reliable, pervasive, long-term continuous monitoring of vital signs. This includes, for example, the development of conductive fabric and weaving technologies [16] or planar-fashionable circuit board technique to directly print patterned electrodes and circuits on fabric [17]. Attempts have also been made to model the human body as a communication channel [18], [19]. In addition to the earlier emerging technologies, the widespread use of mobile devices with high-speed Internet connection and GPS location has greatly extended the possible scope of BSNs.

Among the many applications of BSN, healthcare is likely to be an early adopter of the technology for managing both chronic diseases and acute events. This can potentially change the snapshot nature of conventional monitoring approaches, as the current practice is generally limited to the brief time points and often unrepresentative physiological states such as supine and sedated, or artificially introduced exercise tests. Transient abnormalities cannot always be captured reliably. Important and even life-threatening disorders can go undetected because they occur only infrequently, and much time and effort is wasted in trying to capture an “episode” with controlled monitoring. The provision of “ubiquitous” and “pervasive” monitoring of physical, physiological, and biochemical parameters in any environment without activity restriction and behavior modification is the primary motivation of BSN research. Such an approach has already shown promising results in monitoring patients after surgery, during rehabilitation, as well as for assessing activities of daily living for the elderly. Further development of BSN will continue to pave the way for implementing population-wide, multimodality health records such as the Cardiovascular Health Informatics and Multimodal E-record (CHIME) [20].

The intelligence of the system facilitates the use of physiological monitors as “guardian angels” to provide timely warnings and guidance for a variety of medical conditions or wellness management scenarios. In some situations, a feedback loop can be formed within a BSN that integrates wearable and/or implantable devices. For example, the communication network that forms between a cochlear implant and an external wearable unit that processes speech and sound and transmits control signals to the implant can be considered as a BSN. The wearable unit also supplies energy to the implant, where change of battery is inconvenient. Similar ideas have been adopted for visual implant to restore visual sensations in patients with blindness and retinal degenerations [21]. Continuous real-time control of physiological parameters such as blood glucose and blood pressure is another emerging application of BSN. These systems allow adaptive algorithms to be implemented to avoid hypoglycemia or hypotension. Proactive sensors such as automatic inflatable protective airbag for fall protection have also been developed in BSN research.




Fig. 1. Illustration of the design space of the special issue on body sensor networks.

III. CURRENT ISSUES AND OBSTACLES

Thus far, increasing number of research groups around the world have initiated a range of clinical and ambulatory monitoring projects for a number of health conditions. Examples include:

- postoperative monitoring of patients;
- monitoring of patients with chronic diseases;
- social networking of relatives and peers for monitoring of elderly;
- lifestyle and general well-being monitoring (e.g., to deal with obesity);
- wellness and exercise monitoring;
- monitoring vitals and status of soldiers and firefighters;
- emergency medical care and mass casualty events;
- computer-assisted rehabilitation and therapy; and
- development of new emergency services with prolonged monitoring.

Existing deployment of long-term monitoring has revealed a number of issues that may impede the wider acceptance of BSNs. The problems identified include:

- reliability of short range communication;
- intermittent network coverage in rural areas;
- changes of policy and organizational flaw to regulate payments and reimbursements of physicians for monitoring/consultation;
- insufficient battery life;
- motion artifacts and sensitivity to sensor placements;
- lack of seamless integration with infrastructure network and electronic medical records;
- privacy and security considerations; and
- standardization and interoperability.

IV. SPECIAL ISSUE

Papers in this special issue reflect a maturing trend of the BSN technologies. We received 33 papers in response to the call, ten papers were accepted for the special issue. They are illustrated in Fig. 1 and belong to the following general categories:

- development of new applications of BSNs;
- development of new sensors and sensor interfaces;

- development of wearable actuators for prevention of injuries;
- optimization of sensors; and
- characterization and optimization of wireless communication channels and biochannels.

Development of new applications is presented in the papers by Patel *et al.* and Estudillo-Valderrama *et al.* Patel *et al.* present preliminary results of evaluation of possible use of wireless sensors for automated analysis of activity of Parkinson's patients [22]. This research represents a growing trend of applications aimed at quantitative rather than subjective evaluation of the condition of these patients. In this paper, data from wearable inertial sensors are used to determine motor fluctuations as an indication of the effectiveness and timing of typical medications for Parkinson's patients. The second paper by Estudillo-Valderrama *et al.* outlines the use of inertial sensors for in-home monitoring of a distributed fall detection system [23].

New applications require new sensors or optimization of the existing sensors. Vogel *et al.* present the development of a new vital sign monitoring system placed in the ear of the user [24]. The proposed approach allows embedding of physiological sensors in hearing aid devices or wireless headsets.

Monitoring of multiple physiological parameters leads to more cumbersome systems and decreased convenience for the user. Therefore, an emerging trend is minimization of the number of sensors and evaluation of physiological parameters from the reduced set of sensors. Boyle *et al.* present possible use of the existing cardiac sensor for the estimation of respiration rate [25]. The proposed approach reduces the number of sensors required and enhances user's convenience. However, the current generation of cardiac monitors requires the use of clinical "wet" electrodes that may irritate the skin of most users after prolonged use. Therefore, one of the very active research fields is the implementation of "dry" electrodes with minimum skin irritation or embedding of electrodes in the shirt to improve user's convenience. Yoo *et al.* present a cardiac monitoring system with planar-fashionable circuit board shirt [17].

Existing processors could be optimized to provide application specific tradeoff between the required performance on demand and low-power consumption during regular operation. Raskovic and Giessel present a case for dynamic voltage and frequency scaling to provide higher performance of monitoring systems on-demand [26]. This approach can be used to optimize the existing processors before establishing detailed requirements for application specific hardware.

Increased sophistication of sensors and real-time on-node processing can facilitate the development of new applications that will enable proactive system operation to protect the user. As an example, Tamura presents a wearable system with airbags that could be used to prevent fall injuries of elderly or patients with movement disorders [27].

It is widely recognized that wireless communication consumes a significant portion of available power of monitoring systems. Three papers in this special issue reflect current research efforts in the field of low-power and secure wireless com-

munication. Marinkovic *et al.* present an optimized MAC protocol for energy-efficient communication [28]. Power-efficient security is another important issue of wireless BANs. In this issue, Tan *et al.* present a lightweight-identity-based cryptography for resource-constrained platform for BSNs [29]. Power optimization requires optimization of every element of the system, which includes wireless communication channels. Reusens *et al.* present the comparison of simulations and actual measurements of on-body communication channels to determine optimal configuration and operation [19].

Papers published in this special issue represent a snapshot of the current research and implementations of BSN, as they are maturing to become commercially viable systems. We thank all the authors who have submitted their papers to this special issue. Due to space and topical constraints, many of the quality papers submitted are not able to be included in this special issue. We hope TITB will become a regular forum for reporting, as well as discussing, the latest developments in BSNs.

ACKNOWLEDGMENT

This work was supported in part by the Hong Kong Innovation and Technology Fund (ITF). The authors are grateful to Standard Telecommunication Ltd., Jetfly Technology Ltd., Golden Meditech Company Ltd., Bird International Ltd., and Bright Steps Corporation for their supports to the ITF projects.

EMIL JOVANOVIĆ
Electrical and Computer Engineering Department
University of Alabama in Huntsville
Huntsville, AL 35899 USA

CARMEN C. Y. POON
Department of Electronic Engineering
The Chinese University of Hong Kong
Shatin, Hong Kong

GUANG-ZHONG YANG
Institute of Biomedical Engineering and the
Department of Computing
Imperial College London
London, SW7 U.K.

Y. T. ZHANG
Department of Electronic Engineering
The Chinese University of Hong Kong
Shatin, Hong Kong;
Institute of Biomedical and Health Engineering
Shenzhen Institute of Advanced Technology
Chinese Academy of Sciences (CAS), China;
CAS Key Laboratory for Biomedical Informatics and
Health Engineering, China

REFERENCES

- [1] E. Jovanov, J. Price, D. Raskovic, K. Kavi, T. Martin, and R. Adhami, "Wireless personal area networks in telemedical environment" in *Proc. 3rd IEEE EMBS Inf. Technol. Appl. Biomed. Workshop Int. Telemed. Inf. Soc. (ITAB ITIS 2000)*, Arlington, VA, Nov. 2000, pp. 22–27.
- [2] T. G. Zimmerman, "Personal Area Networks: Near-field intrabody communication," *IBM Syst. J.*, vol. 35, no. 3–4, pp. 609–617, 1996.
- [3] B. Heile, I. Gifford, and T. Siep, "The IEEE P802.15 working group for wireless personal area networks," *IEEE Netw.*, vol. 13, no. 4, pp. 4–5, Jul. 1999.
- [4] K. van Dam, S. Pitchers, and M. Barnard, "From PAN to BAN: Why body area networks," presented at the Wireless World Res. Forum 2, Helsinki, Finland, May 10–11, 2001.
- [5] R. Schmidt, T. Norgall, J. Mörsdorf, J. Bernhard, and T. von der Grün, "Body area network BAN—a key infrastructure element for patient-centered medical applications," *Biomed. Tech. (Berl.)*, vol. 47, pp. 365–368, 2002.
- [6] A. van Halteren, R. Bults, K. Wac, N. Dokovsky, G. Koprnikov, I. Widya, D. Konstantas, V. Jones, and R. Herzog, "Wireless body area networks for healthcare: The MobiHealth project," in *Wearable eHealth Systems for Personalised Health Management: State of the Art and Future Challenges*, vol. 108, Amsterdam, The Netherlands: IOS Press, 2004.
- [7] R. S. H. Istepanian, E. Jovanov, and Y. T. Zhang, "Guest editorial introduction to the special section on M-health: Beyond seamless mobility and global wireless health-care connectivity," *IEEE Trans. Inf. Technol. Biomed.*, vol. 8, no. 4, pp. 405–414, Dec. 2004.
- [8] G. Z. Yang, *Body Sensor Networks*. London, U.K.: Springer-Verlag, 2006.
- [9] D. Panescu, "Implantable neurostimulation devices," *IEEE Eng. Med. Biol. Mag.*, vol. 27, no. 5, pp. 100–113, Sep.–Oct. 2008.
- [10] C. C. Y. Poon, Y. T. Zhang, and S. D. Bao, "A novel biometrics method to secure wireless body area sensor networks for telemedicine and m-Health," *IEEE Commun. Mag.*, vol. 44, no. 4, pp. 73–81, Apr. 2006.
- [11] C. H. Chan, C. C. Y. Poon, R. C. S. Wong, and Y. T. Zhang, "A hybrid body sensor network for continuous and long-term measurement of arterial blood pressure," in *Proc. 4th IEEE-EMBS Int. Summer School Symp. Med. Devices Biosens.*, Cambridge, U.K., Aug. 19–22, 2007, pp. 121–123.
- [12] A. Wong, D. McDonagh, O. Omeni, C. Nunn, M. Silveira, and A. Burdett, "Sensium: An ultra-low-power wireless body sensor network platform: Design & application challenges," *Proc. 31th Int. Conf. IEEE Eng. Med. Biol. Soc.*, Minneapolis, MN, Sep. 2–6, 2009, pp. 6576–6579.
- [13] J. Kwong, Y. K. Ramadass, N. Verma, and A. P. Chandrakasan, "A 65 nm sub- V_t microcontroller with integrated SRAM and switched capacitor DC–DC converter," *IEEE J. Solid State Circuits*, vol. 44, no. 1, pp. 115–126, Jan. 2009.
- [14] V. Leonov, B. Gyselinckx, C. Van Hoof, T. Torfs, F. Yazicioglu, R. J. M. Vullers, and P. Fiorini, "Wearable self-powered wireless devices with thermoelectric energy scavengers," in *Proc. Conf. Smart Syst. Integr.*, Barcelona, Spain, Apr. 9–10, 2008, pp. 217–224.
- [15] M. J. Cooney, V. Svoboda, C. Lau, G. Martina, and S. D. Minter, "Enzyme catalysed biofuel cells," *Energy Environ. Sci.*, vol. 1, pp. 320–337, 2008.
- [16] R. Paradiso, G. Loriga, and N. Taccini, "A wearable health care system based on knitted integrated sensors," *IEEE Trans. Inf. Technol. Biomed.*, vol. 9, no. 3, pp. 337–344, Sep. 2005.
- [17] J. Yoo, L. Yan, S. Lee, H. Kim, and H. J. Yoo, "A wearable ECG acquisition system with compact planar-fashionable circuit board based shirt," in this issue.
- [18] M. S. Wegmueller, A. Kuhn, J. Froehlich, M. Oberle, N. Felber, N. Kuster, and W. Fichtner, "An attempt to model the human body as a communication channel," *IEEE Trans. Biomed. Eng.*, vol. 54, no. 10, pp. 1851–1857, Oct. 2007.
- [19] E. Reusens, W. Joseph, B. Latré, B. Braem, G. Vermeeren, L. Martens, I. Moerman, and C. Blondia, "Characterization of on-body communication channel and energy efficient topology design for wireless body area networks," in this issue.
- [20] Y. T. Zhang, C. C. Y. Poon, and E. MacPherson, "Editorial note on health informatics," *IEEE Trans. Inf. Technol. Biomed.*, vol. 13, no. 3, pp. 281–283, May 2009.
- [21] G. Roessler, T. Laube, C. Brockmann, T. Kirschkamp, B. Mazinani, M. Goertz, C. Koch, I. Krisch, B. Sellhaus, H. K. Trieu, J. Weis, N. Bornfeld, H. Rothgen, A. Messner, W. Mokwa, and P. Walter, "Implantation and explantation of a wireless epiretinal retina implant device: Observations during the EPIRET3 prospective clinical trial," *Invest. Ophthalmol. Vis. Sci.*, vol. 50, no. 6, pp. 3003–3008, Jun. 2009.
- [22] S. Patel, K. Lorincz, R. Hughes, N. Huggins, J. Growdon, D. Standaert, M. Akay, J. Dy, M. Welsh, and P. Bonato, "Monitoring motor fluctuations in patients with Parkinson's disease using wearable sensors," in this issue.
- [23] M. A. Estudillo-Valderrama, L. M. Roa, J. Reina-Tosina, and D. Naranjo-Hernández, "Design and implementation of a distributed fall detection system," in this issue.
- [24] S. Vogel, M. Hulsbusch, T. Hennig, V. Blazek, and S. Leonhardt, "IN-EAR vital signs monitoring using a novel micro-optic reflective sensor," in this issue.
- [25] J. Boyle, N. Bidargaddi, A. Sarela, and M. Karunanithi, "Automatic detection of respiration rate from ambulatory single lead ECG," in this issue.
- [26] D. Raskovic and D. Giessel, "Dynamic voltage and frequency scaling for on-demand performance and availability of biomedical embedded systems," in this issue.
- [27] T. Tamura, "A wearable airbag to prevent fall injuries," in this issue.
- [28] S. J. Marinkovic, E. M. Popovici, C. Spagnol, S. Faul, and W. P. Marnane, "Energy-efficient low duty cycle MAC protocol for wireless body area networks," in this issue.
- [29] C. C. Tan, H. Wang, S. Zhong, and Q. Li, "IBE-lite: A lightweight identity based cryptography for body sensor networks," in this issue.



Emil Jovanov (M'98–SM'04) received the Dipl.Ing., M.Sc., and Ph.D. degrees in electrical and computer engineering from the University of Belgrade, Belgrade, Serbia. From 1984 to 1998, he was with the "Mihajlo Pupin" Research Institute, Belgrade. Between 1994 and 1998, he was an Adjunct Assistant Professor with the University of Belgrade. In 1998, he joined the University of Alabama, Huntsville, where he is currently an Associate Professor with the Electrical and Computer Engineering Department. He is the Originator of the concept of wireless body area networks for health monitoring and a Leader in the field of wearable health monitoring. He has been engaged in developing intelligent sensors for personal health monitoring and mobile computing for more than ten years. He was the Principal Investigator or Coinvestigator on several grants from the National Science Foundation and industry in the field of wireless and sensor networks, wireless intelligent sensors, and wearable health monitors. His current research interests include ubiquitous and mobile computing, biomedical signal processing, and health monitoring. He has authored or coauthored 12 book chapters, 34 journal papers, and 120 conference papers.

He is a member of the Editorial Board of the *International Journal of Telemedicine and Application and Applied Psychophysiology and Biofeedback*.

Dr. Jovanov was the Guest Editor of the IEEE TRANSACTIONS ON INFORMATION TECHNOLOGY IN BIOMEDICINE, SPECIAL ISSUE ON M-HEALTH: BEYOND SEAMLESS MOBILITY AND GLOBAL WIRELESS HEALTH-CARE CONNECTIVITY published in December 2004. He is an Associate Editor of the TRANSACTIONS ON INFORMATION TECHNOLOGY IN BIOMEDICINE and the TRANSACTIONS ON BIOMEDICAL CIRCUITS AND SYSTEMS. He is a member of the IEEE Engineering in Medicine and Biology Society (IEEE-EMBS) Technical Committee on Wearable Biomedical Sensors and Systems and a member of the IEEE Medical Technology Policy Committee.



Carmen C. Y. Poon (M'08) received the B.A.Sc. degree in engineering science (biomedical option) and the M.A.Sc. degree in biomedical engineering from the University of Toronto, Toronto, ON, Canada, the Ph.D. degree from The Chinese University of Hong Kong, Shatin, Hong Kong, where she is currently a Research Assistant Professor. Her current research interests include biosignal processing, biosystem modeling, and development of wearable medical devices and body sensor network for telemedicine and m-Health.

Dr. Poon is an Associate Editor of the *TRANSACTIONS ON INFORMATION TECHNOLOGY IN BIOMEDICINE*.



Guang-Zhong Yang (SM) received the Ph.D. degree in computer science from the Imperial College London, London, U.K.

His current research interests include medical imaging, robotics, and sensing. He is the Director of the Medical Imaging and Robotics with the Institute of Biomedical Engineering, Imperial College London, where he is the Chair of the Imperial College Imaging Sciences Centre and the Founding Director of the Royal Society/Wolfson Medical Image Computing Laboratory.

Prof. Yang is an Associate Editor for the *IEEE TRANSACTIONS ON MEDICAL IMAGING*. He has received several major international awards, including the I. I. Rabi Award from the International Society for Magnetic Resonance in medicine and the Royal Society Research Merit Award in 2001 for medical image computing.



Yuan-Ting Zhang (M'90–SM'93–F'06) received the Master's degree from Shandong University, Jinan, China, and the Ph.D. degree from the University of New Brunswick, Fredericton, NB, Canada, in 1990.

He is the Head of the Division of Biomedical Engineering and the Director of the Joint Research Center for Biomedical Engineering, The Chinese University of Hong Kong, Shatin, Hong Kong, China. He is also the Director of Key Laboratory for Biomedical Informatics and Health Engineering, Chinese Academy of Sciences (CAS), where he is the Director of the Institute of Biomedical and Health Engineering, Shenzhen Institute of Advanced Technology. During 1996–1997 and 2000–2001, he is the Chair with the Biomedical Division, Hong Kong Institution of Engineers. His current research interests include neural engineering, terahertz imaging, and wearable medical devices and body sensor networks, particularly for mobile health. He has authored or coauthored more than 300 scientific articles in the area of biomedical engineering.

Dr. Zhang was the Technical Program Chair of the 20th Annual International Conference in 1998 and the General Conference Chair of the 27th Annual International Conference in 2005. He was elected as an IEEE-EMBS AdCom member in 1999 and was the Vice President (Conferences) in 2000. He was an Associate Editor for the *IEEE TRANSACTIONS ON BIOMEDICAL ENGINEERING* and the *IEEE TRANSACTIONS ON MOBILE COMPUTING*. He was also the Guest Editor of the *IEEE COMMUNICATION MAGAZINE* and the *IEEE TRANSACTIONS ON INFORMATION TECHNOLOGY IN BIOMEDICINE*. He is currently the Editor-in-Chief of the *IEEE Transactions on Information Technology in Biomedicine*. He is an Honorary Advisor of the Hong Kong Medical and Healthcare Device Manufacture Association.